

Finite® P Indigo AF Nitrile Gloves

Finite Indigo AF is a premium powder free nitrile disposable glove compliant to European Standard EN374, providing an effective barrier against oils, acids and other chemicals. It has an AQL rating of 1.5 and complies to Medical Standard EN455 Class 1.

Made from high quality nitrile synthetic rubber, it is extremely soft with increased elasticity, yet still remains a tensile strength comparable to natural rubber. It is ambidextrous and has a unique shape with an improved fit.

The cuff is beaded which adds strength and prevents liquid roll back. Finite P Indigo AF is free of proteins found in natural rubber eliminating protein sensitisation and is free from residual accelerators that are a common cause of allergic dermatitis. It conforms to standard EN1149 for anti-static properties and is powder free reducing the risk of product contamination.



FEATURES

- AQL Rating 1.5
- Powder Free
- Complies to Medical Standard EN455 Class 1
- Beaded Cuff
- Ambidextrous
- 24cm Length
- Antistatic EN1149

It is important to note that the performance of these gloves may be different if mixtures of chemicals are involved.

TYPICALLY USED FOR

- Pharmaceutical Duties
- Electronics
- Inspection
- Laboratory Work
- Local Authority Work
- Hospital & Clinical Duties
- Hospital & Clinical Duties
- Automotive Maintenance & Repair

PRODUCT CODE	SIZE	PACK QTY
602-0098	XS	Box of 100
602-0099	S	Box of 100
602-0100	M	Box of 100
600-0638	L	Box of 100
602-0102	XL	Box of 100

CE0088



DATASHEET

Chemical Resistance guide for Indigo

It is important to note that the performance of these gloves may be different if mixtures of chemicals are involved.

CHEMICAL	EN374 PERFORMANCE LEVEL
Acetonitrile	0
n-Heptane	1
Hydrochloric acid 30%	4
Isopropyl alcohol	1
Sodium hydroxide 40%	6
Toluene	0

EN374 Performance Levels

PERFORMANCE LEVEL	1	2	3	4	5	6
Breakthrough time (mins)	>10	>30	>60	>120 >	240	>480

The information contained here is advisory and is intended to guide and inform only. This information is based on the view on Integrity® Cleanroom in light of all available data and is not subject to any guarantee on our part regarding the performance of a particular glove style. As it is impossible to test every situation encountered and all chemical and mixtures of chemicals, the user is advised to test the particular glove style before and periodically during use.

Statement of Product Conformity

Certified to ISO 9001 & ISO 13485

declare that the new Medical Device and PPE described hereafter



FINITE P INDIGO AF

Non-sterile, powder-free, synthetic nitrile rubber
disposable examination glove

is in conformity with the provisions of Council Directive 93/42/EEC and with National Standards transposing harmonised standards EN455-1, EN455-2, EN455-3, EN455-4 and is self-certified as a Class 1 non-sterile medical device

Freedom from holes: AQL 1.5
Physical properties: 6N minimum

is in conformity with the provisions of Council Directive 89/686/EEC and where such is the case, with National Standards transposing harmonised standards EN374-1:2003, EN374-2:2003, EN374-3:2003 and EN420:2003.

is identical to the PPE which is the subject of EC certificate of conformity No. 3766 issued by SATRA Technology Centre, Wyndham Way, Telford Way, Kettering, Northamptonshire, NN16 BSD, United Kingdom (Notified Body No. 0321).

conforms to the provisions of Council Directive 89/686/EEC Article 11B under the supervision of the Notified Body, LRQA Ltd, Hiramford, Middlemarch Office Village, Siskin Drive, Coventry, CV3 4FJ (Notified Body No. 0088).

MICRO ORGANISM
HAZARDS
EN374-2



CHEMICAL
HAZARDS
EN374-3



Issue No. 5

Model. Finite P Indigo AF

Done at Enfield on 27/01/14

DATASHEET

CE 0088



Complex Design

User Information

This product does not contain natural rubber latex.
This product does not contain chemical accelerators.

Glove Care

Storage: Store below 40°C in dry conditions away from direct sunlight and heat.
You are advised to retain this packaging for reference

Range Available: sizes and codes

Sizes	5.5	6.5	7.5	8.5	9.5
Code	MFNP100/00	MFNP100/01	MFNP100/02	MFNP100/03	MFNP100/04

Food Contact



Finite P Indigo AF have been tested in accordance with EN1186. They are suitable for use with all food categories in situations of short term, repeat contact.

Issue No. 5

Model: Finite P Indigo AF

Done at Enfield on 27/01/14